

GeoAnalytical Laboratories, Inc.

2300 Maryann Dr. Turlock, CA 95380 Phone (209) 669-0100 Fax (209) 593-2212
email: info@geoanalyticallab.com

Report # A4C1910

Project: SCVWD Penitencia Clearwell Tank

Date: 03/25/14

Bay Area Coating Consultants
P.O. Box 867
Denair, CA 95316

Date Rec'd: 03/19/14

CERTIFICATE OF ANALYSIS

Sampler: Ed Darrimon

Sample ID: Shell South Side


Lab ID	Sample Date / Time	RL	Method	Analyte	Result	Units	Notes	Started
A4C1910-01	3/13/2014 13:00	1.0	6020	Lead	570	mg/kg		3/24/14 11:00
		0.5		Zinc	2000	mg/kg		


Sample ID: Shell North Side

Lab ID	Sample Date / Time	RL	Method	Analyte	Result	Units	Notes	Started
A4C1910-02	3/13/2014 13:00	1.0	6020	Lead	560	mg/kg		3/24/14 11:00
		0.5		Zinc	790	mg/kg		

Sample ID: Center Roof

Lab ID	Sample Date / Time	RL	Method	Analyte	Result	Units	Notes	Started
A4C1910-03	3/13/2014 13:00	1.0	6020	Lead	7.3	mg/kg		3/24/14 11:00
		0.5		Zinc	1000	mg/kg		


Cedric Gregory
Chemist


Donna Keller
Laboratory Director

GeoAnalytical Laboratories, Inc.

2300 Maryann Dr. Turlock, CA 95380 Phone (209) 669-0100 Fax (209) 593-2212
email: info@geoanalyticallab.com

Report # A4C1910

Project: SCVWD Penitencia Clearwell Tank

Date: 03/25/14

Bay Area Coating Consultants
P.O. Box 867
Denair, CA 95316

Date Rec'd: 03/19/14

Inorganic Chemistry - Quality Control GeoAnalytical Laboratories, Inc.

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch A000834 - 3050B										
Prepared & Analyzed: 03/24/14										
Blank (A000834-BLK1)										
Lead	ND	1.0	mg/kg							
Zinc	ND	0.5	"							
LCS (A000834-BS1)										
Lead	12.8	1.0	mg/kg	12.50		103	80-120			
Zinc	12.8	0.5	"	12.50		102	80-120			
LCS Dup (A000834-BSD1)										
Lead	12.5	1.0	mg/kg	12.50		100	80-120	2	20	
Zinc	12.5	0.5	"	12.50		100	80-120	2	20	
Matrix Spike (A000834-MS1)										
Source: A4C1402-01										
Lead	17.7	1.0	mg/kg	12.50	3.95	110	80-120			
Zinc	73.8	0.5	"	12.50	71.7	16	80-120			QM-02
Matrix Spike Dup (A000834-MSD1)										
Source: A4C1402-01										
Lead	17.4	1.0	mg/kg	12.50	3.95	108	80-120	1	20	
Zinc	76.2	0.5	"	12.50	71.7	36	80-120	3	20	QM-02

GeoAnalytical Laboratories, Inc.

2300 Maryann Dr. Turlock, CA 95380 Phone (209) 669-0100 Fax (209) 593-2212
email: info@geoanalyticallab.com

Report # A4C1910

Project: SCVWD Penitencia Clearwell Tank

Date: 03/25/14

Bay Area Coating Consultants
P.O. Box 867
Denair, CA 95316

Date Rec'd: 03/19/14

Notes and Definitions

- QM-02 The RPD and/or percent recovery for this QC spike sample cannot be accurately calculated due to the high concentration of analyte inherent in the sample.
- ND Analyte NOT DETECTED at or above the reporting limit
- RPD Relative Percent Difference
- RL Reporting Limit
- NA Not Applicable